

	INSTRUMENT DATA SHEET	DOCUMENT NO. 100-0189 Rev. 17
SAGE RIO THERMAL MASS FLOW METER	GAS MASS FLOW	SAGE SRX SPECIFICATIONS REMOTE STYLE ATEX INSERTION MASS FLOW METER

GENERAL INFORMATION

STYLE:	Remote Insertion Mass Flow Meter
SENSOR:	Two reference grade platinum RTD clad in 316SS (1.4401) sheath
MATERIAL:	Wetted metal components: 316SS
POWER:	24VDC or 115/230VAC optional
POWER DISSIPATION:	<2.5 W (for 24VDC Models)
ELECTRONICS:	Remote-Style Microprocessor based (Hybrid-Digital)
PROBE ENCLOSURE:	Class 1, Div. 1 & 2, Group B, C, D, Zone 1
TRANSMITTER ENCLOSURE:	Zone 1, Group IIB+H2, IP66
DISPLAY:	High contrast photo-emissive OLED graphical display (Flow Rate, Totalizer, Temperature)
TURNDOWN:	100 to 1
RESOLUTION:	1000 to 1
LOW END SENSITIVITY:	25 mm/sec (5 SFPM)
FIELD CALIBRATION CHECK:	Yes - Digital system allows raw signal validation in milliwatts (In-Situ Calibration Check)
COMMUNICATIONS:	Modbus® compliant RS485 RTU or optional HART® communications
APPROVALS:	ATEX Zone I Ex II 2 G Ex d IIB+H2 T6 Gb Note: T6 Rating is suitable for gases with ignition temperatures as low as 185°F (85°C) CSA: Can/CSA C22.2 UL: UL 61010-1 3rd edition
FIELD RECONFIGURABLE:	SageCom and Modbus required
FLOW ACCURACY:	+/- 0.5% of Full Scale +/- 1% of reading (Enhanced accuracy optionally available with limited turn-down)
FLOW REPEATABILITY:	0.2%
RESPONSE TIME:	1 second time constant
GAS TEMPERATURE RANGE:	Standard -40° to 200°F (93°C), Optional to 300°F (149°C) and 450°F (232°C)
GAS PRESSURE:	34 Bar (500 PSIG) (If higher pressure needed, contact Sage)
FLOW OUTPUT:	4 to 20 mA for Rate
TOTALIZER:	24VDC pulse for Totalized value
TEMPERATURE:	On Display/Output with Modbus®
AMBIENT TEMPERATURE:	-40° to 122°F (50°C) for ATEX Rating
PROBE STYLE / LENGTH:	12.7mm (1/2") OD Probe Lengths 152mm (6") to 914mm (36")
CABLE LENGTH:	7.6 meters (25') Standard - max length 305 meters (1000')
FLOW CONDITIONING:	Captive Flow Conditioners available upon request with meter purchase
NOTES:	This unit is provided with ATEX approved cord grips

